

Fieldbus	Specif/Config file	D5	MF5	F5	F6	F6 Fieldbus Software	Hilscher Firmware	GSD-EDS Config Files	Hilscher Hardware Ref	Ref Kit ATEQ
PROFIBUS	-	> 1.21E	> 1.000y1	> 1.18L			V01.058 (19/12/03)	HIL_7501.GSD (09/03/2000)	COM-DPS	9600297
PROFIBUS	-	>3.10c5 (2012)	>3.10c5 (2012)	>3.10h5 (2012)	> 1.060 (14/10/14)	> 1.05	2.4.5.7	HIL_7501.GSD (09/03/2000)	COMX100	9600297B
DEVICENET	-	> 1.21E	> 1.000y1	> 1.18L			V01.043 (09/12/03)	COMDNS.EDS (28/03/2000)	COM-DPS	9600662
DEVICENET	-	>3.10c5 (2012)	>3.10c5 (2012)	>3.10h5 (2012)	> 1.060 (14/10/14)	> 1.05	2.3.3.2	COMDNS.EDS (28/03/2000)	COMX100	9600660
PROFINET	V2.3	>3.10g11 (07/07/14)	>=3.10g5 (07/07/14)	>=3.10L9 (07/07/14)	> 1.060 (14/10/14)	> 1.05	3.4.44.0 -- (DIM 3)	GSDML-V2.3-HILSCHER-COMX 100XX-RE PNS-20130806.xml	COMX100	9600659
PROFINET	V2.31	>=3.10j4	>=3.10i2	>=3.10p2	>=1.318	> 2.103	3.5.43-3.5.X -- (DIM24)	GSDML-V2.31-HILSCHER-COMX 51XX-RE PNS-20140519.xml	COMX51	9600590
ETHERNET IP	V1.1	>3.10d5 (2012)	>3.10c5 (2012)	>3.10h5 (2012)		> 1.05	2.5.16.4	HILSCHER COMX 100XX-RE EIS V1.1.EDS (2011)	COMX100	9600661
ETHERNET IP	V1.1	>=3.10g11 (07/07/14)	>=3.10g5 (07/07/14)	>=3.10L9 (07/07/14)	> 1.060 (14/10/14)	> 1.05	2.7.13.0	HILSCHER COMX 100XX-RE EIS V1.1.EDS (21/02/12)	COMX100	9600661
ETHERCAT	V4.6	3.10j3	3.10i1	3.10p1	>=1.318		4.6.0.1	Hilscher COMX 51XX RE ECS V4.6.X.xml (13/10/2016)	COMX51	

Note How to select Config file
 1- Check board (COMX 100/COMX 51) (Menu Service, Fieldbus)
 2- Check Hilscher Firmware (Menu Service, Fieldbus)
 3- Select Config. File
 4- Verify if Ateq Software supported, if not upgrade it

Ex : Profinet COMX 100, Firmware : 3.4.44 -- GSDML-V2.3-HILSCHER-COMX 100XX-RE PNS-20130806.xml

Warning For PLC STEP7 V5.5 SP 4 GSD For Profinet COMX 51 (Use GSDMH V2.13)